




A92150Mu03 Polyclonal Antibody to Collagen Type VI Alpha 1 (COL6a1) Instruction manual

FOR IN VITRO USE AND RESEARCH USE ONLY
NOT FOR USE IN CLINICAL DIAGNOSTIC PROCEDURES

5th Edition (Revised in January, 2013)

[PRODUCT INFORMATION]



Sample: Recombinant mouse COL6a1

[IMMUNOGEN INFORMATION]

Immunogen: Recombinant mouse COL6a1 (Asn82~Phe212) expressed in

E.coli.

Molecular Weight: 16.0kDa

USCN accession No.: P92150Mu03

Sequence: The target protein is fused with N-terminal His-Tag and its sequence

is listed below.

MGHHHHHHSGSEF- NLVWNAGAL HYSDEVEIIR GLTRMPSGRD ELKASVDAVK
YFGKGTYTDC AIKKGLEELL IGGSHLKENK YLIVVTDGHP LEGYKEPCGG LEDAVNEAKH
I GIKVESVAI TPDHI EPRI S IIATDHTYRR NE

Unique product Superb quality Client favorite Nicest service @ ISO9001:2008; @ ISO13485:2003; $C \in S$





[ANTIBODY SPEFICITY]

Anti COL6a1 is a rabbit polyclonal antibody raised against mouse COL6a1. This antibody has been selected for its ability to recognize mouse COL6a1 in immunohistochemical staining and western blotting, non cross-reactive with other members of the family.

[APPLICATIONS]

Western blotting: 1:100-400

Immunocytochemistry in formalin fixed cells: 1:100-500

Immunohistochemistry in formalin fixed frozen section: 1:100-500

Immunohistochemistry in paraffin section: 1:50-200

Optimal working dilutions must be determined by end user.

[CONTENTS]

Form & Buffer: Supplied as solution form in PBS, pH 7.4, containing 0.02%Na₃N, 50% glycerol.

[QUALITY CONTROL]

Content: The quality control contains recombinant mouse COL6a1

(Asn82~Phe212) disposed in loading buffer.

Usage: 10uL per well when 3,3'-Diaminobenzidine(DAB) as the substrate.

5uL per well when used in enhanced chemilumescent (ECL).

Note: The quality control is specifically manufactured as the positive control. Not used for other purposes.

Loading buffer: 100mM Tris(PH 8.8), 2%SDS, 200mM NaCl, 50% glycerol, BPB 0.01%, Na₃N 0.02%.

[STORAGE]

Store at 4°C for frequent use. Stored at -20°C to -80°C in a manual defrost freezer for one year without detectable loss of activity. Avoid repeated freeze-thaw cycles.