

MAA541Hu28 Monoclonal Antibody to Brain Natriuretic Peptide (BNP) Organism Species: *Homo sapiens (Human) Instruction manual* 

FOR RESEARCH USE ONLY

NOT FOR USE IN CLINICAL DIAGNOSTIC PROCEDURES

12th Edition (Revised in Aug, 2016)

# Cond-Clone Corp.

## [PROPERTIES]

Source: Monoclonal antibody preparation Host: Hybridoma culture supernatant Antibody isotype: IgG2b Kappa Purification: Protein A + Protein G affinity chromatography Clone number: C3 Traits: Liquid Concentration: 1mg/ml UOM: 100µg(100µl) Cross Reactivity: Mouse; Rat; Pig Applications: WB; IHC; ICC; IP. [IMMUNOGEN ] Immunogen: Synthetic Peptide, BNP conjugated to OVA.Target peptide sequence: SPKMVQGSGCFGRKMDRI

Accession No.: CPA541Hu21

## [APPLICATIONS]

Western blotting: 0.5-2µg/mL;

Immunohistochemistry: 5-20µg/mL;

Immunocytochemistry: 5-20µg/mL;

Optimal working dilutions must be determined by end user.

#### [FORMULATION]

**Form & Buffer:** Supplied as solution form in PBS, pH7.4, containing 0.02% NaN3, 50% glycerol.

#### [ STORAGE AND STABILITY ]

Storage: Avoid repeated freeze/thaw cycles.


Store at 4°C for frequent use.

Aliquot and store at -20°C for 12 months.

## Control Cloud-Clone Corp.

**Stability Test:** The thermal stability is described by the loss rate. The loss rate was determined by accelerated thermal degradation test, that is, incubate the protein at 37°C for 48h, and no obvious degradation and precipitation were observed. The loss rate is less than 5% within the expiration date under appropriate storage condition.

### [IDENTIFICATION]



Western Blot; Sample: Porcine Skeletal muscle lysate; Primary Ab: 2µg/ml Mouse Anti-Human BNP Antibody Second Ab: 0.2µg/mL HRP-Linked Caprine Anti-Mouse IgG Polyclonal Antibody (Catalog: SAA544Mu19)




Figure. DAB staining on IHC-P; Samples: Human Liver Tissue.




Figure. DAB staining on IHC-P; Samples: Human Stomach Tissue.

## [<u>IMPORTANT NOTE</u>]

The kit is designed for research use only, we will not be responsible for any issue if the kit was used in clinical diagnostic or any other procedures.