

RPL617Hu01 50µg Recombinant Insulin Like Growth Factor 2 mRNA Binding Protein 3 (IGF2BP3) Organism Species: *Homo sapiens (Human)* Instruction manual

FOR RESEARCH USE ONLY

NOT FOR USE IN CLINICAL DIAGNOSTIC PROCEDURES

13th Edition (Revised in Aug, 2023)

# Coud-Clone Corp.

# [PROPERTIES]

Source: Prokaryotic expression

Host: E.coli

Residues: Asn2~lle343

Tags: N-terminal His Tag

Subcellular Location: Nucleus, Cytoplasm

**Purity:** > 95%

Traits: Freeze-dried powder

Buffer formulation: 20mM Tris, 150mM NaCl, pH8.0, containing 1mM EDTA, 1mM DTT,

0.01% Sarcosyl, 5% Trehalose and Proclin300.

Original Concentration: 200µg/mL

Applications: Positive Control; Immunogen; SDS-PAGE; WB.

(May be suitable for use in other assays to be determined by the end user.)

Predicted isoelectric point: 8.5

Predicted Molecular Mass: 41.6kDa

Accurate Molecular Mass: 42kDa as determined by SDS-PAGE reducing conditions.

#### [<u>USAGE</u>]

Reconstitute in 20mM Tris, 150mM NaCl (pH8.0) to a concentration of 0.1-1.0 mg/mL. Do not vortex.

#### [STORAGE AND STABILITY]

Storage: Avoid repeated freeze/thaw cycles.

Store at 2-8°C for one month.

Aliquot and store at -80°C for 12 months.


**Stability Test:** The thermal stability is described by the loss rate. The loss rate was determined by accelerated thermal degradation test, that is, incubate the protein at 37°C for 48h, and no obvious degradation and precipitation were observed. The loss rate is less than 5% within the expiration date under appropriate storage condition.



### [<u>SEQUENCE</u>]

NKLYIGNLS	ENAAPSDLES	IFKDAKIPVS	GPFLVKTGYA	<b>FVDCPDESWA</b>
LKAIEALSGK	IELHGKPIEV	EHSVPKRQRI	RKLQIRNIPP	HLQWEVLDSL
LVQYGVVESC	EQVNTDSETA	VVNVTYSSKD	QARQALDKLN	GFQLENFTLK
VAYIPDEMAA	QQNPLQQPRG	RRGLGQRGSS	RQGSPGSVSK	QKPCDLPLRL
LVPTQFVGAI	IGKEGATIRN	ITKQTQSKID	VHRKENAGAA	EKSITILSTP
EGTSAACKSI	LEIMHKEAQD	IKFTEEIPLK	ILAHNNFVGR	LIGKEGRNLK
KIEQDTDTKI	TISPLQELTL	YNPERTITVK	GNVETCAKAE	EEI

## [IDENTIFICATION]



#### [IMPORTANT NOTE]

The kit is designed for research use only, we will not be responsible for any issue if the kit was used in clinical diagnostic or any other procedures.